

ANALYTICAL REPORT

Client: Lake Sinissippi Improvement District
 Attn: Dave Behl
 PO Box 89
 Hustisford, WI 53034

NLS Project: 363094
NLS Customer: 107350
Phone: 815 219 9270

Project: 16-777

Hwy S Bridge NLS ID: 1244471

COC: 215337:1 Matrix: SW
 Collected: 03/17/21 00:00 Received: 03/18/21

Parameter	Result	Units	Dilution	LOD	LOQ	Analyzed	Method	Lab
BOD-5 day	3.1	mg/L	1			03/18/21	5210 B-2001	721026460
Nitrogen, ammonia as N (unfiltered)	[0.066]	mg/L	1	0.027	0.090	03/19/21	4500-NH3 G-1997	721026460
Nitrogen, NO2 + NO3 as N (unfiltered)	1.4	mg/L	1	0.040	0.14	03/25/21	4500-NO3 F-2000	721026460
Nitrogen, Kjeldahl as N (unfiltered)	1.2	mg/L	1	0.20	0.68	03/24/21	EPA 351.2, Rev 2	721026460
Phosphorus, tot. as P	0.12	mg/L	1	0.0060	0.020	03/24/21	4500-P E-1999	721026460
Solids, tot. susp. (TSS)	9.3	mg/L	.794	1.6*		03/19/21	2540 D-1997	721026460

Hustisford Dam NLS ID: 1244472

COC: 215337:2 Matrix: SW
 Collected: 03/17/21 00:00 Received: 03/18/21

Parameter	Result	Units	Dilution	LOD	LOQ	Analyzed	Method	Lab
BOD-5 day	6.4	mg/L	1			03/18/21	5210 B-2001	721026460
Nitrogen, ammonia as N (unfiltered)	ND	mg/L	1	0.027	0.090	03/19/21	4500-NH3 G-1997	721026460
Nitrogen, NO2 + NO3 as N (unfiltered)	1.5	mg/L	1	0.040	0.14	03/25/21	4500-NO3 F-2000	721026460
Nitrogen, Kjeldahl as N (unfiltered)	1.7	mg/L	1	0.20	0.68	03/24/21	EPA 351.2, Rev 2	721026460
Phosphorus, tot. as P	0.14	mg/L	1	0.0060	0.020	03/24/21	4500-P E-1999	721026460
Solids, tot. susp. (TSS)	25	mg/L	1	2.0*		03/19/21	2540 D-1997	721026460

Dead Creek NLS ID: 1244473

COC: 215337:3 Matrix: SW
 Collected: 03/17/21 00:00 Received: 03/18/21

Parameter	Result	Units	Dilution	LOD	LOQ	Analyzed	Method	Lab
BOD-5 day	<2	mg/L	1			03/18/21	5210 B-2001	721026460
Nitrogen, ammonia as N (unfiltered)	0.12	mg/L	1	0.027	0.090	03/19/21	4500-NH3 G-1997	721026460
Nitrogen, NO2 + NO3 as N (unfiltered)	2.6	mg/L	1	0.040	0.14	03/25/21	4500-NO3 F-2000	721026460
Nitrogen, Kjeldahl as N (unfiltered)	1.1	mg/L	1	0.20	0.68	03/24/21	EPA 351.2, Rev 2	721026460
Phosphorus, tot. as P	0.099	mg/L	1	0.0060	0.020	03/24/21	4500-P E-1999	721026460
Solids, tot. susp. (TSS)	7.7	mg/L	.794	1.6*		03/19/21	2540 D-1997	721026460

Hwy S Springs NLS ID: 1244474

COC: 215337:4 Matrix: SW
 Collected: 03/17/21 00:00 Received: 03/18/21

Parameter	Result	Units	Dilution	LOD	LOQ	Analyzed	Method	Lab
BOD-5 day	<2	mg/L	1			03/18/21	5210 B-2001	721026460
Nitrogen, ammonia as N (unfiltered)	[0.028]	mg/L	1	0.027	0.090	03/19/21	4500-NH3 G-1997	721026460
Nitrogen, NO2 + NO3 as N (unfiltered)	5.0	mg/L	1	0.040	0.14	03/25/21	4500-NO3 F-2000	721026460
Nitrogen, Kjeldahl as N (unfiltered)	[0.48]	mg/L	1	0.20	0.68	03/25/21	EPA 351.2, Rev 2	721026460
Phosphorus, tot. as P	0.065	mg/L	1	0.0060	0.020	03/24/21	4500-P E-1999	721026460
Solids, tot. susp. (TSS)	2.7	mg/L	.769	1.5*		03/19/21	2540 D-1997	721026460

NORTHERN LAKE SERVICE, INC.
 Analytical Laboratory and Environmental Services
 400 North Lake Avenue - Crandon, WI 54520
 Ph: (715)-478-2777 Fax: (715)-478-3060

ANALYTICAL REPORT

WDNR Laboratory ID No. 721026460
 WDATCP Laboratory Certification No. 105-330
 EPA Laboratory ID No. WI00034

Printed: 03/26/21 Page 2 of 2

Client: Lake Sinissippi Improvement District
 Attn: Dave Behl
 PO Box 89
 Hustisford, WI 53034

NLS Project: 363094
 NLS Customer: 107350
 Phone: 815 219 9270

Project: 16-777

Horicon Golf NLS ID: 1244475

COC: 215337:5 Matrix: SW

Collected: 03/17/21 00:00 Received: 03/18/21

Parameter	Result	Units	Dilution	LOD	LOQ	Analyzed	Method	Lab
BOD-5 day	<2	mg/L	1			03/18/21	5210 B-2001	721026460
Nitrogen, ammonia as N (unfiltered)	[0.039]	mg/L	1	0.027	0.090	03/19/21	4500-NH3 G-1997	721026460
Nitrogen, NO2 + NO3 as N (unfiltered)	1.8	mg/L	1	0.040	0.14	03/25/21	4500-NO3 F-2000	721026460
Nitrogen, Kjeldahl as N (unfiltered)	[0.46]	mg/L	1	0.20	0.68	03/25/21	EPA 351.2, Rev 2	721026460
Phosphorus, tot. as P	0.060	mg/L	1	0.0060	0.020	03/24/21	4500-P E-1999	721026460
Solids, tot. susp. (TSS)	4.8	mg/L	.746	1.5*		03/19/21	2540 D-1997	721026460

Values in brackets represent results greater than or equal to the LOD but less than the LOQ and are within a region of "Less-Certain Quantitation". Results greater than or equal to the LOQ are considered to be in the region of "Certain Quantitation". LOD and/or LOQ tagged with an asterisk(*) are considered Reporting Limits. All LOD/LOQs adjusted to reflect dilution and/or solids content.

ND = Not Detected (< LOD) LOD = Limit of Detection LOQ = Limit of Quantitation NA = Not Applicable
 DWB = Dry Weight Basis %DWB = (mg/kg DWB) / 10000 1000 ug/L = 1 mg/L
 MCL = Maximum Contaminant Levels for Drinking Water Samples. Shaded results indicate >MCL.

Reviewed by:



Authorized by:
 R. T. Krueger
 President